

Art.-Nr.: KG2306
Ch.-B.: 812LO
Inhalt: 5 x 1 ml
Verw. bis: 2010-05

ASO Standard is a human serum material. Refer to the table below for the appropriate assigned ASO value.

Cat No.	Name	ASO Value (IU/ml)
LO 795	ASO	486
LO 7935	ASO	486
LO 3835	ASO	486
LO 3998	ASO 2	484
LO 3999	ASO 2	484
LO 2877	ASO	486
LO 2888	ASO 2	484

The Standard is for use in the determination of Anti-Streptolysin-O. Calibration of ASO Standard has been performed at BGT by latex enhanced immunoturbidimetry with reference to a material standardised against the appropriate International Reference Preparation.

RECONSTITUTION AND STABILITY

The Standard is a liquid preparation which is supplied ready for use and is stable up to the expiry date when capped and stored at +2 to +8°C in the absence of contamination. Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

SAFETY PRECAUTIONS AND WARNINGS

Human source material from which this product has been derived has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE.

FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

For in vitro diagnostic use only, do not pipette by mouth, exercise the normal precautions required for handling laboratory reagents.

This material contains Sodium Azide. Avoid ingestion or contact with skin or mucous membranes. In case of skin contact, flush affected area with copious amounts of water. In case of contact with eyes or if ingested, seek immediate medical attention.

Sodium Azide reacts with lead and copper plumbing, to form potentially explosive azides.

When disposing of such reagents flush with large volumes of water to prevent azide build up.

Exposed metal surfaces should be cleaned with 10% sodium hydroxide.

18 Nov '08