

Art.-Nr.: KG2609
Ch.-B.: 123GR
Inhalt: 10 x 5 ml
Verw. bis: 2011-11

Preparation

Open the vial very carefully, avoiding any loss of material and reconstitute in exactly 5 mls of freshly distilled water at +20 - +25°C. Replace the rubber stopper, close vial and leave to stand for 30 minutes before use.

Stability

Stable for 1 day at +2 - +8°C, or 8 hours at +15 - +25°C.

Calibration value (U/l)	58
--------------------------------	----

Safety

This control is manufactured from bovine serum. Human additives which include prostatic acid phosphatase and prostate specific antigen have been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests. However as no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious disease.

For IN VITRO diagnostic use only.

Revised 30 Sep '08