

**Art.-Nr.:** KG1381  
**Ch.-B.:** 019MC  
**Inhalt:** 5 x 2 ml  
**Verw. bis:** 2009-10

**INTENDED USE**

This product is intended for *in-vitro* diagnostic use in the control of lactate/glucose/ammonia/glutamate on RX Daytona clinical chemistry systems.

**SAFETY PRECAUTIONS AND WARNINGS**

For *in-vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.  
Health and Safety Data Sheets are available on request.

**STORAGE AND STABILITY**

**Opened:** Store refrigerated (2°C to 8°C). Stable to expiration date if kept capped in original containers and free of contamination.

**Unopened:** Store refrigerated (2°C to 8°C). Stable to expiration date printed on the individual vials.

**PREPARATION FOR USE**

1. Mix by inversion at least 5 times to ensure homogeneity prior to use.
2. Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

**MATERIALS PROVIDED**

Multi Control Level 3 (5 x 2 ml)

**MATERIALS REQUIRED BUT NOT PROVIDED**

None.

**ASSIGNED VALUES**

A value (mean) has been assigned to this lot of Control at BGT Laboratories Ltd.

<b>ANALYTE</b>	<b>TARGET VALUE</b>	<b>RANGE</b>
<b>LACTATE</b>	44.33 MMOL/L	37.68 - 50.98
<b>GLUCOSE</b>	24.26 MMOL/L	20.62 - 27.90
<b>AMMONIA</b>	8076 UMOL/L	6865 - 9287
<b>GLUTAMATE</b>	8460 UMOL/L	7191 - 9729
	8.46 MMOL/L	7.19 - 9.73

04 Dec '08