



Reference Values provided by DYMIND

CALIBRATOR

LOT PLUS0424


2024-03-12

2024-05-05

| Applicable Instruments | Parameter | Unit | Reference Value |
|--|-----------|--------------------|-----------------|
| DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0) Hem-o-test 3600 | WBC | $\times 10^9/L$ | 9.27 |
| | RBC | $\times 10^{12}/L$ | 4.71 |
| | HGB | g/L | 135 |
| | MCV | fL | 93.3 |
| | PLT | $\times 10^9/L$ | 238 |
| DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher) | WBC | $\times 10^9/L$ | 9.17 |
| | RBC | $\times 10^{12}/L$ | 4.52 |
| | HGB | g/L | 135 |
| | MCV | fL | 86.4 |
| | PLT | $\times 10^9/L$ | 239 |
| DH20, DH22 (Technical File Version B1.0 or higher) | WBC | $\times 10^9/L$ | 9.35 |
| | RBC | $\times 10^{12}/L$ | 4.58 |
| | HGB | g/L | 136 |
| | MCV | fL | 90.8 |
| | PLT | $\times 10^9/L$ | 237 |
| DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher) | WBC | $\times 10^9/L$ | 9.20 |
| | RBC | $\times 10^{12}/L$ | 4.52 |
| | HGB | g/L | 130 |
| | MCV | fL | 93.6 |
| | PLT | $\times 10^9/L$ | 236 |

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.



Reference Values provided by DYMIND

CALIBRATOR

LOT PLUS0424

2024-03-12

2024-05-05

| Applicable Instruments | Parameter | Unit | Reference Value |
|---|-----------|--------------------|-----------------|
| DH36X (Technical File Version A1.0 or higher) | WBC | $\times 10^9/L$ | 9.63 |
| | RBC | $\times 10^{12}/L$ | 4.63 |
| | HGB | g/L | 136 |
| | MCV | fL | 89.9 |
| | PLT | $\times 10^9/L$ | 234 |
| DP-H10 Vet (Technical File Version A1.0 or higher) | WBC | $\times 10^9/L$ | 9.19 |
| | RBC | $\times 10^{12}/L$ | 4.52 |
| | HGB | g/L | 140 |
| | MCV | fL | 93.6 |
| | PLT | $\times 10^9/L$ | 236 |

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.